

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Monty Ross on 4 December 2007.

The application has been amended as follows:

DB
1/2
In claim 113, after "forward^{END} of the spring" in the last line of the claim, "wherein a portion of the needle holder extends forwardly of any portion of the barrel" is added.

Claim 115 is cancelled.

The Abstract is replaced with the following paragraph:

A syringe assembly having a retractable needle, the syringe assembly being rendered unusable after a single injection and having a hollow syringe body, a retraction mechanism with a spring disposed in the front portion of the syringe and an inner head, a continuous retainer member surrounding the inner head, and a bridging portion disposed between the continuous retainer member and the inner head, wherein the bridging portion couples the continuous retainer member and the inner head to form a fluid seal between a fluid passageway and the barrel prior to retraction, and a plunger reciprocally disposed inside the barrel and forming a variable chamber between the plunger and the needle holder prior to and during retraction, wherein the continuous